

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 08/01/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:28
INTERNAL STANDARD	VERIFIED	18:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:29
BLANK TEST	0.00	18:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31

