

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 02/02/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14

