

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/01/2025

BFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:35
INTERNAL STANDARD	VERIFIED	23:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:36
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:37
BLANK TEST	0.00	23:38

