

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 08/28/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:21
BLANK TEST	0.00	07:22

