

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 08/06/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:09
INTERNAL STANDARD	VERIFIED	00:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13

