

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/19/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:43
INTERNAL STANDARD	VERIFIED	05:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:44
BLANK TEST	0.00	05:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:45
BLANK TEST	0.00	05:46

