

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/26/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00

