

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/15/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:55
INTERNAL STANDARD	VERIFIED	17:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59

