

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 141906

Date: 02/05/2025

OPERATOR NAME:  
 THOMAS R BRONSON  
 Permit #: 53522  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	VERIFIED	22:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:54
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:57
BLANK TEST	0.00	22:58

