

# SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 10/08/2025

OPERATOR NAME:  
ELDAR MUJIC  
Permit #: 54489  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 22:42 |
| INTERNAL STANDARD       | VERIFIED | 22:42 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 22:42 |
| BLANK TEST              | 0.00     | 22:43 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 22:43 |
| BLANK TEST              | 0.00     | 22:44 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 22:44 |
| BLANK TEST              | 0.00     | 22:45 |

