

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/17/2025

OPERATOR NAME:
MARK F PACIELLO
Permit #: 53019
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:11
INTERNAL STANDARD	VERIFIED	06:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:12
BLANK TEST	0.00	06:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:14
BLANK TEST	0.00	06:15

