

# SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 10/22/2025

OPERATOR NAME:  
THOMAS R BRONSON  
Permit #: 53522  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 00:22 |
| INTERNAL STANDARD       | VERIFIED | 00:22 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 00:22 |
| BLANK TEST              | 0.00     | 00:23 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 00:23 |
| BLANK TEST              | 0.00     | 00:24 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 00:24 |
| BLANK TEST              | 0.00     | 00:25 |

