

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/24/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:48
INTERNAL STANDARD	VERIFIED	16:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:50
BLANK TEST	0.00	16:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:51
BLANK TEST	0.00	16:52

