

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/08/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:21
INTERNAL STANDARD	VERIFIED	18:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:22
BLANK TEST	0.00	18:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:23
BLANK TEST	0.00	18:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:25

