

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 01/15/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:51
INTERNAL STANDARD	VERIFIED	21:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:51
BLANK TEST	0.00	21:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:53
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:54
BLANK TEST	0.00	21:55

