

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 02/19/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:59
INTERNAL STANDARD	VERIFIED	17:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:01
BLANK TEST	0.00	18:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:02
BLANK TEST	0.00	18:03

