

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/16/2025

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:15
INTERNAL STANDARD	VERIFIED	17:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:15
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:17
BLANK TEST	0.00	17:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:18
BLANK TEST	0.00	17:19

