

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

QFO

Date: 02/12/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:51
INTERNAL STANDARD	VERIFIED	20:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:52
BLANK TEST	0.00	20:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:53
BLANK TEST	0.00	20:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:54
BLANK TEST	0.00	20:55

