

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/03/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:57
INTERNAL STANDARD	VERIFIED	23:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:58
BLANK TEST	0.00	23:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01

