

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/01/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:34
INTERNAL STANDARD	VERIFIED	22:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37

