

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/03/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:24
INTERNAL STANDARD	VERIFIED	18:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:26
BLANK TEST	0.00	18:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:27
BLANK TEST	0.00	18:28

