

SOLUTION CHANGE

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 141906

Date: 05/07/2026

BFO

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:11
INTERNAL STANDARD	VERIFIED	15:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:13
BLANK TEST	0.00	15:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:14
BLANK TEST	0.00	15:15

