

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/24/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:28
INTERNAL STANDARD	VERIFIED	20:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:28
BLANK TEST	0.00	20:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:30
BLANK TEST	0.00	20:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32

