

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/13/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:59
INTERNAL STANDARD	VERIFIED	23:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:02

