

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/22/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:08
INTERNAL STANDARD	VERIFIED	23:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:11
BLANK TEST	0.00	23:12

