

**SUPERVISOR TEST**

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906  
-----

Date: 04/08/2026

OPERATOR NAME:  
THOMAS R BRONSON  
Permit #: 53522  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44

