

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 03/11/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34

