

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 07/17/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48

