

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/22/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:57
INTERNAL STANDARD	VERIFIED	18:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:59
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:00
BLANK TEST	0.00	19:01

