

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/07/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:59
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:01
BLANK TEST	0.00	23:02

