

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/29/2026

OPERATOR NAME:
MARC A PERRY
Permit #: 55276
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:39
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:40
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42

BFO

