

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 02/26/2026

OPERATOR NAME:  
THOMAS R BRONSON  
Permit #: 53522  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:33
INTERNAL STANDARD	VERIFIED	15:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:34
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:36
BLANK TEST	0.00	15:37

