

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 02/05/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:08
INTERNAL STANDARD	VERIFIED	18:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:09
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:11
BLANK TEST	0.00	18:11

