

SUPERVISOR TEST

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 141906

Date: 02/11/2026

OPERATOR NAME:
 MARC A PERRY
 Permit #: 55276
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	VERIFIED	22:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:54
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57

BFD

