

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/16/2026

OPERATOR NAME:
THOMAS R BRONSON
Permit #: 53522
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:46
INTERNAL STANDARD	VERIFIED	17:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49

