

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/22/2026

BFC

OPERATOR NAME:  
THOMAS R BRONSON  
Permit #: 53522  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:12
INTERNAL STANDARD	VERIFIED	05:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:13
BLANK TEST	0.00	05:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:14
BLANK TEST	0.00	05:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:15
BLANK TEST	0.00	05:16

