

SUPERVISOR TEST

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906  
-----

Date: 05/13/2026

OPERATOR NAME:  
THOMAS R BRONSON  
Permit #: 53522  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:59
BLANK TEST	0.00	23:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:01
BLANK TEST	0.00	23:02

