

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 09/16/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:57
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:00
BLANK TEST	0.00	01:00

