

SOLUTION CHANGE

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 12/01/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54

