

SUPERVISOR TEST

ONEIDA CO.
 SHERIFF'S DEPT
 DATAMASTER dmt: 123906

Date: 08/18/2021

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:16
INTERNAL STANDARD	VERIFIED	21:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:17
BLANK TEST	0.00	21:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:18
BLANK TEST	0.00	21:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:19
BLANK TEST	0.00	21:20

