

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/13/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:27
INTERNAL STANDARD	VERIFIED	19:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:30
BLANK TEST	0.00	19:31

