

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 10/28/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:51
INTERNAL STANDARD	VERIFIED	04:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:51
BLANK TEST	0.00	04:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:53
BLANK TEST	0.00	04:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:54
BLANK TEST	0.00	04:55

