

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 11/03/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:58
INTERNAL STANDARD	VERIFIED	03:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:59
BLANK TEST	0.00	04:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:00
BLANK TEST	0.00	04:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:01
BLANK TEST	0.00	04:02

