

SUPERVISOR TEST

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

CFO

Date: 12/22/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:13
INTERNAL STANDARD	VERIFIED	23:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:15
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:17

