

Oneida County Sheriff's Department  
Breath Test Instrument  
Camden Field Office

900A - 9304314-001  
DATAMASTER-~~900152~~

123906

Date	Operator	Defendant	Chemical Test	Purge Analysis	Instrument Number	Arresting Officer
8/19/20	Grabeldiner	weekly	.10 x 3	34.00- 34.01	123906	—
8/26/20	RUTENGE	WEEKLY	.10 x 3	34.01 x 3	123906	—
9/2/20	Osika	Weekly	.10 x 3	34.00 34.01 x 2	123906	—
9/9/20	Karcic	Weekly	.10 x 3	34.00 34.01 x 2	123906	—
9/16/20	Grabeldiner	weekly	.10 x 3	34.00- 34.01	123906	—
9/25/20	Karcic	weekly	.10 x 3	34.01 x 3	123906	—
9/30/20	Grabeldiner	SOLUTION CHANGE	.09 x 3	34.01 x 3	123906	—
10-4-20	Karcic	Joseph W. Zimmer	.18	34.01	123906	Z. Przybyla
10-7-20	Cornish	Weekly	.09 x 3	34.01 x 3	123906	—
10/15/20	Karcic	Weekly	.09 x 3	34.00 34.01 x 2	123906	—
10/22/20	LATIC	Weekly	.09 x 3	34.01 x 3	123906	—
10/29/20	Karcic	Weekly	.09 x 3	34.00 x 2 34.01	123906	—
11/1/20	LATIC	Poplawski, Edward R.	.20	34.01	123906	Matthew Hoffman
11/02/2020	Villenbeck	—	.09	34.00/34.01	123906	—
11/5/20	Karcic	Weekly	.09 x 3	34.01 34.00 x 2	123906	—
11/12/20	Karcic	Weekly	.09 x 3	34.01 x 3	123906	—
11/16/20	Grabeldiner	Weekly	.09 x 3	34.00 34.01	123906	—
11/26/20	Karcic	Weekly	.09 x 3	34.01 x 2 34.00	123906	—
12/1/20	Karcic	Solution change	.09 x 3	34.01 x 3	123906	—
12/2/20	Grabeldiner	weekly	.09 x 3	34.00- 34.01	123906	—
12/9/20	Karcic	Weekly	.09 x 3	34.01 34.00 x 2	123906	20430
12/17/20	LATIC	Weekly	.09 x 3	34.00 34.01 x 2	123906	20430
12/24/20	Karcic	Weekly	.09 x 3	34.00 34.01 x 2	123906	—

Oneida County Sheriff's Department  
Breath Test Instrument  
Camden Field Office

900A - 9304314-001  
DATAMASTER-900152

123906

Date	Operator	Defendant	Chemical Test	Purge Analysis	Instrument Number	Arresting Officer
12/15/20	MACK	weekly				
1/1/21	MARCANO	<del>solution change</del> <sup>N.M</sup>	.10	34.01	123906	
1/1/21	MARCANO	solution change	.09x3	34.00 34.01x2	123906	
1/6/21	MARCANO	WEEKLY	.09x3	34.00x3	123906	
1/8/21	MARCANO	NATHAN MOSLEY	REFUSED	REFUSED	123906	KARCIC
1/14/21	MACK	weekly	.10	34.01	123906	
1/20/21	MARCANO	WEEKLY	.09x3	34.01x3	123906	
1/27/21	MACK	weekly	.10	34.01	123906	
2/1/21	MACK	solution change	.09x3	34.00 x2 34.01	123906	
2/3/21	MACK	Solution change	.10x2 .09	34.01 x2 34.0	123906	
2/10/21	MACK	WEEKLY	.09 x2 .10	34.01 34.00 x2	123906	
2/18/21	MARCANO	WEEKLY	.09 x 2 .10	34.00 x 2 34.01	123906	
2/24/21	MARCANO	WEEKLY	.09 x 3	34.00x3	123906	
3/1/21	MACK	Solution change	.09 x 2 .10	34.00 34.01 x2	123906	
3/3/21	Karcic	weekly	.09x3	34.00 34.01x2	123906	
3/10/21	MACK	weekly	.09x3	34.01 x2 34.00 x2	123906	
3/27/21	MACK	weekly	.09x3	34.01 x2 34.0	123906	
3/31	Karcic	weekly	.09x3	34.00x3	123906	
4/2/21	MARCANO	SOLUTION CHANGE	.10 x 3	34.01 x 2 34.00	123906	
4/7/21	MARCANO	WEEKLY	.10x3	34.01 x2 34.00	123906	
4/14/21	Karcic	Weekly	.09x3	34.01x2 34.00	123906	
4/21/21	MARCANO	WEEKLY	.09 x 3	34.00x3	123906	
4/28/21	Karcic	weekly	.10x3	34.01x3	123906	
4/28/21	Karcic	Weekly	.09x3	34.01x3	123906	

Oneida County Sheriff's Department  
Breath Test Instrument  
Camden Field Office

900A - 9304314-001  
DATAMASTER-~~900152~~  
123906

Date	Operator	Defendant	Chemical Test	Purge Analysis	Instrument Number	Arresting Officer
5/1	Cornish	Solution Change	.10	34.00 x 2 34.01	123906	—
5/5/21	MARCANO	WEEKLY	.10	34.01 x 3	123906	—
5/12/21	MARCANO	WEEKLY	.10	34.01 x 3	123906	—
5/19/21	MARCANO	WEEKLY	.09	34.01 x 3	123906	—
5/27/21	Karcic	weekly	.09 x 3	34.01 x 3	123906	—
6/2/21	Karcic	Solution Change <sup>+ Lot # Change</sup>	.09 x 3	34.01 x 2 34.00	123906	—
6/2/21	Karcic	weekly	.10 x 3	34.00 x 3	123906	—
6/10/21	Eilers	weekly	.10 x 3	34.01 x 3	123906	—
6/16/21	Przybyla	Weekly	.09 x 3	34.00 x 1 34.01 x 2	123906	—
6/23/21	Eilers	Weekly	.09 x 3	34.01 x 2 34.00	123906	—
6/30/21	MARCANO	WEEKLY	.09 x 3	34.01 x 2 34.00	123906	—
7/2/21	Przybyla	Weekly / Bottle change	.10 x 3	34.01 x 2 34.00	123906	—
7-7-21	Eilers	weekly	.10 x 3	34.01 x 3	123906	—
7/15/21	Przybyla	Weekly	.09 x 1	34.01 x 3	123906	—
7/16/21	Eilers	Weekly	.09 x 2 .10 x 1	34.01 x 3	123906	—
7/22/21	Przybyla	Weekly	.09 x 3	34.01 x 3	123906	—
8/1/21	Osika	Solution Change	.10 x 3	34.01 x 3	123906	—
8-4-21	Eilers	Weekly	.10 x 2 .09 x 1	34.01 x 3	123906	—
8/12/21	Przybyla	Weekly	.09 x 3	34.00 x 1 34.01 x 2	123906	—
8-18-21	Eilers	Weekly	.09 x 3	34.01 x 3	123906	—
8-22-21	Foster, Morgan	Chamberlain, Mark S.	.14	34.01	123906	Cornish
8-25-21	Przybyla	Weekly	.09 x 3	34.01 x 2 34.00 x 1	123906	—



Oneida County Sheriff's Department  
Breath Test Instrument  
Camden Field Office

900A - 9304314-001  
DATAMASTER-~~930152~~

123906

Date	Operator	Defendant	Chemical Test	Purge Analysis	Instrument Number	Arresting Officer
9-1-21	Eilers	Solution Change	.09	34.01x3	123906	—
9-16-21	Eilers	DMT Setup & test	.10x2 .09x1	34.01x3	123906	Weekly
9/24/21	HAYNES	COMO, FRANK, J.	.15		123906	CORNISH
9/29/21	Eilers	Weekly	.09x2 .10x1	34.01	123906	Weekly
10/4/21	Eilers	Solution Change	.10	34.01x2 34.00	123906	21280
10/11/21	Millenbeck	Aldridge, Paula L	Refused	Refused	123906	McDaniels
10/13/21	Eilers	Weekly	.10	34.01	123906	21280
10/20/21	HAYNES	weekly test	.10	34.01	123906	21280
10/27/21	Eilers	Weekly	.10 .09x2	34.01	123906	21280
11/03/21	Eilers	Weekly/Bottle change	.10x3	34.01x2 34.00	123906	21280
11/11/21	Quinn	weekly	.10x3	34.01x2 34.00x1	123906	21280
11/24/21	Eilers	weekly	.09x3	34.01x3	123906	21280
12/1/21	Eilers	Solution Change	.10x3	34.01x2 34.00	123906	21390
12/8/21	Eilers	Weekly	.10x3	34.01	123906	21390
12/16/21	Cornish	Weekly	.09 .10x2	34.01x2 34.00	123906	21390
12/22/21	Eilers	Weekly	.09x3	34.01x3	123906	21390
12/30/21	Przybylo	Weekly	.09x3	34.01x3	123906	21390
01/01/22	Eilers	Bottle Change	.10x3	34.01x3	123906	21390
1/19/22	Eilers	Weekly	.10x3	34.01x2 34.00	123906	21390
1/26/22	Osika	Weekly	.10x3	34.00x2 34.01	123906	21390
1/27/22	HAYNES	Weekly	.10x2 .09	34.01x2 34.00	123906	21390
2/1/22	Osika	Solution Change	.10x3	34.01x3	<del>123906</del> 21440	21440
2/9/22	Osika	Weekly	.10x3	34.01x3	123906	21440