

SOLUTION CHANGE

CFO

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 02/01/2022

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:06
INTERNAL STANDARD	VERIFIED	20:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:08
BLANK TEST	0.00	20:09

