

SOLUTION CHANGE

LFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 03/01/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:13
INTERNAL STANDARD	VERIFIED	23:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:15
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:16
BLANK TEST	0.00	23:17

