

SOLUTION CHANGE

CFO

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 04/01/2022

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:05
INTERNAL STANDARD	VERIFIED	23:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:05
BLANK TEST	0.00	23:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:07
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:08
BLANK TEST	0.00	23:09

