

SOLUTION CHANGE

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

CFO

Date: 07/31/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:54
INTERNAL STANDARD	VERIFIED	23:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:54
BLANK TEST	0.00	23:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:56
BLANK TEST	0.00	23:57

