

SOLUTION CHANGE

CFD

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 08/01/2022

OPERATOR NAME:
CHARLES C HAYNES
Permit #: 32834
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:08
INTERNAL STANDARD	VERIFIED	20:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:08
BLANK TEST	0.00	20:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:10
BLANK TEST	0.00	20:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12

